

Test Report No. : CPSA/210257058-CA58515
Company : S Ram & M Ram Resources Berhad
 Suite B-3A-9, Megan Avenue II
 12 Jalan Yap Kwan Seng
 Kuala Lumpur, W.P 50450

TEST REPORT

Sample Description : Nitrile Examination Gloves
 Brand Name : WALLETZ4U
 Style : M Size
 Lot No : 160221-02
 Batch : WZ179PN0002SRJR
 Colour : Blue
 Manufacturer : JR Engineering and Medical Technologies (M) Sdn Bhd
 Production Date : Feb 2021
 Country of Origin : Malaysia
 Size : M
 Quantity Tested : 200 pieces
 Test Conducted : Freedom from holes
 Test Method : EN455 Part 1:2020
 Testing Period : 18 February 2021 - 08 March 2021

Based on submitted samples, the following results obtained:-

Acceptable Quality Limit (AQL) : 1.5 Accept : 7 Found : 2

Result : Within AQL

SIGNED FOR AND ON BEHALF OF
 SGS (MALAYSIA) SDN BHD



CHEE TUCK CHOON
 SECTION HEAD
 IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

Test Report No. : CPSA/210257058-CA58515
Company : S Ram & M Ram Resources Berhad
 Suite B-3A-9, Megan Avenue II
 12 Jalan Yap Kwan Seng
 Kuala Lumpur, W.P 50450

TEST REPORT

Sample Description : Nitrile Examination Gloves
 Brand Name : WALLETZ4U
 Style : M Size
 Lot No : 160221-02
 Batch : WZ179PN0002SRJR
 Colour : Blue
 Manufacturer : JR Engineering and Medical Technologies (M) Sdn Bhd
 Production Date : Feb 2021
 Country of Origin : Malaysia
 Size : M
 Quantity Tested : 13 pieces
 Test Conducted : Dimensions
 Test Method : EN 455 Part 2:2015
 Testing Period : 18 February 2021 - 08 March 2021

Based on submitted samples, the following results obtained :-

Size	M	M	M	M	M	M	M	M	M	M	M	M	M	Median
Width Median: 95±10mm	94	96	96	96	96	96	94	96	95	95	94	96	95	96
Length Median: ≥ 240mm	254	255	256	254	260	250	250	255	250	255	253	255	251	254

SIGNED FOR AND ON BEHALF OF
 SGS (MALAYSIA) SDN BHD



CHEE TUCK CHOON
 SECTION HEAD
 IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

Test Report No. : CPSA/210257058-CA58515
Company : S Ram & M Ram Resources Berhad
Suite B-3A-9, Megan Avenue II
12 Jalan Yap Kwan Seng
Kuala Lumpur, W.P 50450

TEST REPORT

Sample Description : Nitrile Examination Gloves
 Brand Name : WALLETZ4U
 Style : M Size
 Lot No : 160221-02
 Batch : WZ179PN0002SRJR
 Colour : Blue
 Manufacturer : JR Engineering and Medical Technologies (M) Sdn Bhd
 Production Date : Feb 2021
 Country of Origin : Malaysia
 Size : M
 Quantity Tested : 13 pieces per test
 Test Conducted : Force at Break During Shelf Life and After Challenge
 Test Method : EN 455 Part 2:2015
 Ageing : 70 ± 2 Deg C for 168 hrs
 Testing Period : 18 February 2021 - 08 March 2021

SIZE	SAMPLE NO.	Force at Break, N	
		<u>BEFORE AGING</u>	<u>AFTER AGING</u>
M	1	6.5	5.8
	2	5.8	6.0
	3	6.3	7.1
	4	5.7	7.1
	5	6.1	6.4
	6	6.1	6.3
	7	5.9	7.0
	8	7.3	6.4
	9	5.8	6.2
	10	6.6	7.7
	11	6.1	6.2
	12	6.1	6.0
	13	5.8	5.9
Median		6.1	6.3
Requirement		≥ 6.0	≥ 6.0

SIGNED FOR AND ON BEHALF OF
 SGS (MALAYSIA) SDN BHD


 CHEE TUCK CHOON
 SECTION HEAD
 IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

Test Report No. : CPSA/210257058-CA58515
Company : S Ram & M Ram Resources Berhad
 Suite B-3A-9, Megan Avenue II
 12 Jalan Yap Kwan Seng
 Kuala Lumpur, W.P 50450

TEST REPORT

Sample Description : Nitrile Examination Gloves
 Brand Name : WALLETZ4U
 Style : M Size
 Lot No : 160221-02
 Batch : WZ179PN0002SRJR
 Colour : Blue
 Manufacturer : JR Engineering and Medical Technologies (M) Sdn Bhd
 Production Date : Feb 2021
 Country of Origin : Malaysia
 Size : M
 Quantity Tested : 5 pieces
 Test Conducted : Powder Content
 Test Method : EN455 Part 3:2015
 Testing Period : 18 February 2021 - 08 March 2021

On testing the samples, the following results were obtained:-

SIZE

Average Powder Mass per Glove

M

0.28 mg

SIGNED FOR AND ON BEHALF OF
 SGS (MALAYSIA) SDN BHD



CHEE TUCK CHOON
 SECTION HEAD
 IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

Test Report No. : CPSA/210257058-CA58515
Company : S Ram & M Ram Resources Berhad
Suite B-3A-9, Megan Avenue II
12 Jalan Yap Kwan Seng
Kuala Lumpur, W.P 50450

TEST REPORT

Sample Description : Nitrile Examination Gloves
 Brand Name : WALLETZ4U
 Style : M Size
 Lot No : 160221-02
 Batch : WZ179PN0002SRJR
 Colour : Blue
 Manufacturer : JR Engineering and Medical Technologies (M) Sdn Bhd
 Production Date : Feb 2021
 Country of Origin : Malaysia
 Size : M
 Quantity Tested : 8 pieces
 Test Conducted : Aqueous Extractable Protein Content
 Test Method : EN455 Part 3:2015
 Testing Period : 18 February 2021 - 08 March 2021

EXTRACTABLE PROTEIN CONTENT

Analysis had been carried out as per your request. We report the following results:

SIZE	AQUEOUS EXTRACTABLE PROTEIN CONTENT, µg/g
M	NOT DETECTED

Note: Detection limit = 10 µg/g

SIGNED FOR AND ON BEHALF OF
 SGS (MALAYSIA) SDN BHD

**** End of Report ****


 CHEE TUCK CHOON
 SECTION HEAD
 IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.